# Certificate of Conformity

# VWR INOCULATING LOOPS, FLEXIBLE, 10µl, STERILE, IN BAGS OF 10 PCS

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

*VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.*

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **612-9357**

Lot No: **F04400**

Expiry Date: **30th Nov 2015**

Sterility: **By gamma irradiation at 25-35 KGY**
**Procedure compliance with ISO 11137**

**Goods manufactured according to ISO 13485 system.**

*This document has been produced electronically and is valid without a signature*

Date of issue: **December 2012**