

## Certificate of Conformity

# VWR SEROLOGICAL PIPETTES, STANDARD LINE, STERILE

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

*VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.*

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **612-3702**

Lot No: **100415**

| Capacity (ml) | Division (ml) | Colour code | Packed            | Pk         |
|---------------|---------------|-------------|-------------------|------------|
| <b>5</b>      | <b>0.10</b>   | <b>Blue</b> | <b>Individual</b> | <b>300</b> |

Material: **100% Virgin High Clarity Polystyrene**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications.**

Accuracy: **Pipettes are accurate  $\pm 3\%$  at full volume**

Sterility: **Products labeled sterile meets a minimum requirement of  $10^{-6}$  SAL.**

Non-Pyrogenic: **The acceptance level for product is 0.5EU/ml or less than 20 EU/device.  
LAL Gel Clot Method**

Cytotoxicity: **Testing is conducted to quality all material resins using ISO 10993 standards  
and have been shown to be non-toxic**

*This document has been produced electronically and is valid without a signature*

Date of Issue: **July 2013**