



Material	83607.290
Material description	Conductivity standard 1,413 $\mu\text{S}/\text{cm}$ (25°C; 0.01 mol/l; KCl)
Grade	calibration solution for conductivity

Characteristics	Specifications
Titer	0.00995 --> 0.01005 mol/l
Conductivity (25°C)	1.39 --> 1.43 mS/cm

Signature

We certify that this batch conforms to the specifications listed above.

This document has been produced electronically and is valid without a signature.

Anja Vanhalle, Head of Laboratory - Haasrode
 VWR International bvba; Geldenaaksebaan 464; BE-3001
 Leuven; Belgium

For Professional use in Laboratory or Manufacturing. Not for use as an Active Pharmaceutical Ingredient or Food or Animal Feed. Suitability and intended use of the product remains the responsibility of the user.