Certificate of Conformity

VWR SOFT LOOP 1 UL STERILE, IN BAG OF 10 PCS

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **612-9355**

Lot No: **MV6200**

Expiry Date: **31st Oct 2012**

Sterility: **By gamma irradiation at 25-35 KGY**
Procedure compliant to ISO 11137
Tested with 10^6 *B. pumilis* control tube: No growth after 7 days at 33-35°C

Goods manufactured according to ISO 13485 system.

This document has been produced electronically and is valid without a signature

VWR International bvba/sprl, Researchpark Haasrode 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
http://www.vwr.com

Date of issue: **May 2010**